

Certificate of Analysis

Product H50V DETERGENT, RFID
Product No 1440106 (1.9 L)
Lot No 25062047M3
Manufacturing Date 2025-07-09
Expiration Date 2027-07-08

Test	Specification
Appearance	Detergent should be transparent liquid without precipitation, particle or floccules
pH	9.0 – 10.5
Blank	WBC $\leq 0.2 \times 10^9$ /L RBC $\leq 0.02 \times 10^{12}$ /L HGB ≤ 1 g/L PLT $\leq 10 \times 10^9$ /L

Actual values meet specifications.

Date: 2025-07-24



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

