

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22507-41	2025-11-27
Normal	22507-42	2025-11-27
High	22507-43	2025-12-01

MEDONIC

CONTROL		L	
LOT		22507-41	
Medonic M-series/M32		Mean ²	± Range ³
RBC	10 ¹² /L	2.24	0.12
MCV	fL	65.3	5.0
RDW	%	15.1	4.0
Hct	%	14.6	2.5
Plt	10 ⁹ /L	80	15
MPV	fL	10.8	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.4	0.3
MCH	pg	24.1	1.5
MCHC	g/dL	37.0	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	42.4	7.0
Mid	%	6.4	5.0
Gran	%	51.2	7.0

CONTROL		N	
LOT		22507-42	
		Mean ²	± Range ³
		4.15	0.18
		79.6	5.0
		14.7	4.0
		33.0	3.0
		24.0	30
		11.0	1.5
		8.4	0.6
		11.6	0.4
		28.0	1.5
		35.2	3.0
		3.8	0.6
		0.6	0.5
		4.0	1.0
		45.8	5.0
		7.0	5.0
		47.2	5.0

CONTROL		H	
LOT		22507-43	
		Mean ²	± Range ³
		4.94	0.22
		90.4	5.0
		13.5	4.0
		44.7	3.5
		54.0	60
		11.2	1.5
		20.3	1.8
		15.6	0.5
		31.6	1.5
		34.9	3.0
		12.2	1.5
		1.2	1.0
		6.9	2.5
		60.1	5.0
		5.7	5.0
		34.2	5.0

1
bg Трайност на отворен флакон 14 дни (bg)
hr Stabilität offener Flasche 14 Tage (hr)
cs Stabilita otevřených ampulí 14 dní (cs)
da Åben rør stabilitet 14 dage (da)
et Stabiilsus avatud viaalis 14 päevade (et)
fr Stabilité en flacon ouvert 14 jours (fr)
de Stabilität geöffneten Flaschen 14 Tage (de)
el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it Stabilità della fiala aperta 14 giorni (it)
lv Atvērt flakonu stabilitāti 14 dienas (lv)
lt Stabilumas atidarius buteliuką 14 dienas (lt)
no Åpen rør/glass stabilitet 14 dager (no)
pl Trwałość po otwarciu fiolek 14 dni (pl)
pt Estabilidade após abertura do frasco 14 dias (pt)
ro Termenul de valabilitate al fiolei deschise 14 zile (ro)
ru Стабильность вскрытого флакона 14 дней (ru)
sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl Stabilita otvorene injekčneje liekovky 14 dni (sl)
es Estabilidad de la cápsula abierta 14 días (es)
sv Hållbarhet för öppnad flaska 14 dagar (sv)
tr Açık şişe stabilitesi 14 gün (tr)

2
Sredno
Sredstvo
Průměr
Middelværdi
Keskmine
Moyenne
Mittelwert
Μέση
Media
Nozīmē
Vidurkis
Middelverdi
Średnia
Média
Medie
Среднее
Srednja vrednost
Priemer
Povprečno
Media
Medelvärde
Ortalama

3
Обхват
Raspon
Rozsah
Område
Vahemik
Intervalle
Bereich
Αναμενόμενο εύρος α
Intervallo
Diapazons
Intervalas
Område
Zakres
Intervalo
Interval
диапазон
Opseg
Rozsah
Območje
Intervalo
Intervall
Arahğı

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
Lot: 22507-41
Prod.: Con-Diff Low
ID: 2250741+
Exp.: 27-Nov-2025

Instr.: M-series
Lot: 22507-42
Prod.: Con-Diff Norm
ID: 2250742+
Exp.: 27-Nov-2025

Instr.: M-series
Lot: 22507-43
Prod.: Con-Diff High
ID: 2250743+
Exp.: 1-Dec-2025



3.

Boule Con-Diff

CONTROL**IVD**

Control	Lot	
Tri-Level	--	--
Low	22507-41	2025-11-27
Normal	22507-42	2025-11-27
High	22507-43	2025-12-01

Open vial stability 14 days ¹**SWELAB**

		CONTROL	L
		LOT	22507-41
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.25	0.12
MCV	fL	67.1	5.0
RDW	%	14.2	4.0
Hct	%	15.1	2.5
Plt	10 ⁹ /L	81	15
MPV	fL	11.5	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.4	1.5
MCHC	g/dL	36.4	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	47.3	7.0
Mid	%	4.1	5.0
Gran	%	48.6	7.0

		CONTROL	N
		LOT	22507-42
		Mean ²	± Range ³
		4.15	0.18
		81.2	5.0
		13.3	4.0
		33.7	3.0
		245	30
		11.7	1.5
		8.3	0.6
		11.6	0.4
		28.0	1.5
		34.4	3.0
		4.2	0.6
		0.3	0.5
		3.8	1.0
		49.9	5.0
		4.1	5.0
		46.0	5.0

		CONTROL	H
		LOT	22507-43
		Mean ²	± Range ³
		4.91	0.22
		92.4	5.0
		12.1	4.0
		45.4	3.5
		554	60
		11.9	1.5
		20.1	1.8
		15.5	0.5
		31.6	1.5
		34.1	3.0
		12.7	1.5
		0.6	1.0
		6.8	2.5
		63.0	5.0
		3.1	5.0
		33.9	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dni (hr)
cs	Stabilita otevřených ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διάρθρωση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiolek 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvoreneje injekčneje liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

Sredno
Sredstvo
Průměr
Middelverdi
Keskmine
Moyenne
Mittelwert
Μέση
Media
Nožimē
Vidurkis
Middelverdi
Średnia
Média
Medie
Среднее
Srednja vrednost
Priemer
Povprečno
Media
Medelvärde
Ortalama

3

Обхват
Raspon
Rozsah
Område
Vahemik
Intervalle
Bereich
Αναμενόμενο εύρος α
Intervallo
Diapazons
Intervalas
Område
Zakres
Intervalo
Interval
диапазон
Opseg
Rozsah
Območje
Intervalo
Intervall
Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa

Lot: 22507-41

Prod.: Con-Diff Low

ID: 2250741+

Exp.: 27-Nov-2025

Instr.: Swelab Alfa

Lot: 22507-42

Prod.: Con-Diff Norm

ID: 2250742+

Exp.: 27-Nov-2025

Instr.: Swelab Alfa

Lot: 22507-43

Prod.: Con-Diff High

ID: 2250743+

Exp.: 1-Dec-2025

