a Boule Diagnostics, AB company



CERTIFICATE OF ANALYSIS

Product: Boule Cleaning Kit

Kit Lot No: 2275

Product Part #: 501-036

Kit Exp. Date: 2027-05-31

Part Number: 503-085			
Boule Detergent Cleaner Lot # 22	272		
Manufacturing Date: 2025-05-15		Quality Testing Date: 2025-05-19	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 - 8.40	8.27	Pass

Part Number: 503-086			
Boule Hypochlorite (2%) Cl	eaner Lot # 2273		
Manufacturing Date: 2025-05-13		Quality Testing Date: 2025-05-16	
Parameter	Specification	Result	Pass/Fail
pH (units):	11.80 - 13.00	12.93	Pass

Part Number: 503-087				
Boule Enzymatic Cleaner Lot #2	274			
Manufacturing Date: 2025-05-14	Quality Testing Date: 2025-05-19			
Parameter	Specification	Result	Pass/Fail	
pH (units):	7.60 - 8.40	8.06	Pass	

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Quality and Regulatory

Date