

Certificate of Analysis

Product MEDONIC M51-D DILUENT (20L)
Product No 1504510
Lot No 2025051901
Manufacturing Date 2025-05-19
Expiration Date 2027-05-18

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2025-08-07



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

