

# Certificate of Analysis

**Product** Quintus 5-part Lyse 5L  
**Product No** 1504292  
**Lot No** 2506-329  
**Manufacturing Date** 2025-06-16  
**Expiration Date** 2026-12-15

| Test                                    | Specification |
|-----------------------------------------|---------------|
| pH                                      | 2,7 – 3,0     |
| Osmolality (mOsm/kg)                    | 35 – 44       |
| Conductivity (µS/cm)                    | 1117 – 1541   |
| Background WBC ( x 10 <sup>12</sup> /L) | ≤ 0.1         |

Actual values meet specification.

Date: 2025-08-28



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

