


Certificate of Analysis

Product H50V LYSE, RFID
Product No 1440110 (200 mL)
Lot No 25070876M1
Manufacturing Date 2025-07-21
Expiration Date 2027-07-20

Test	Specification
Appearance	Clear liquid, no precipitate particles or flocs
pH	2.00 – 4.50
Absorbance $\lambda_{\max}$ (nm)	530 – 550
Blank	WBC $\leq 0.3 \times 10^9$ /L HGB ≤ 2 g/L
WBC	a) Has two population peaks b) Meet the peak form and position mark requirements.

Actual values meet specifications.

Date: 2025-08-06



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

