a Boule Diagnostics, AB company



1998 FLORIDA

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Lyse Lot No: 2506-301

Product Part #: 501- 291 Exp. Date: 2026-12-31

Manufacturing Date: 2025-06-18

Quality Testing Date: 2025-07-24

Parameter	Specification	Result	Pass/Fail
pH (units):	2.7 – 3.0	2.9	Pass
Osmolality (mOsm/kg):	35 – 44	42	Pass
Conductivity (µS/cm):	1,117 – 1,541	1,236	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Regulatory & Quality Assurance

Date