a Boule Diagnostics, AB company



CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Stopper Lot No: 2506-202

Product Part #: 501-350 Exp. Date: 2028-05-31

Manufacturing Date: 2025-06-23

Quality Testing Date: 2025-07-24

Parameter	Specification	Result	Pass/Fail
pH (units):	10.10 – 11.20	11.01	Pass
Osmolality (mOsm/kg):	830 – 860	835	Pass
Conductivity (mS/cm):	42.7 – 43.3	42.7	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Regulatory & Quality Assurance

Date

1998 FLORIDA