

# Certificate of Analysis

**Product** Exigo System Cleaner, RFID (1.9 L)  
**Product No** 1504503  
**Lot No** 2509-357  
**Manufacturing Date** 2025-09-03  
**Expiration Date** 2028-09-02

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 7,00 – 8,00   |
| Conductivity (mS/cm)        | 15,25 – 17,00 |
| Particles (PLT/RBC channel) | ≤ 30          |

Actual values meet specifications.

Date: 2025-09-08



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

