

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	22509-1K	2026-01-30
Low	22509-11	2026-01-30
Normal	22509-12	2026-01-30
High	22509-13	2026-02-03

MEDONIC

CONTROL		L	
LOT		22509-11	
Medonic M-series/M32		Mean ²	± Range ³
RBC	10 ¹² /L	2.25	0.12
MCV	fL	67.8	5.0
RDW	%	15.5	4.0
Hct	%	15.3	2.5
Plt	10 ⁹ /L	80	15
MPV	fL	11.6	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	24.9	1.5
MCHC	g/dL	36.6	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.8	0.8
Lymph	%	43.7	7.0
Mid	%	5.4	5.0
Gran	%	50.9	7.0

CONTROL		N	
LOT		22509-12	
		Mean ²	± Range ³
		4.15	0.18
		81.0	5.0
		14.1	4.0
		33.6	3.0
		236	30
		11.9	1.5
		8.6	0.6
		11.8	0.4
		28.4	1.5
		35.1	3.0
		4.0	0.6
		0.6	0.5
		4.0	1.0
		46.0	5.0
		6.5	5.0
		47.5	5.0

CONTROL		H	
LOT		22509-13	
		Mean ²	± Range ³
		5.09	0.22
		92.9	5.0
		13.2	4.0
		47.3	3.5
		502	60
		12.1	1.5
		20.6	1.8
		16.2	0.5
		31.8	1.5
		34.2	3.0
		12.5	1.5
		1.2	1.0
		6.9	2.5
		61.0	5.0
		5.6	5.0
		33.4	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dni (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud viaalis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διάρκεια δειγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorene injekčne liekovke 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

Sredno
Sredstvo
Průměr
Middelværdi
Keskmine
Moyenne
Mittelwert
Μέση
Media
Nozīmē
Vidurkis
Middelverdi
Średnia
Média
Medie
Среднее
Srednja vrednost
Priemer
Povprečno
Media
Medelvärde
Ortalama

3

Обхват
Raspon
Rozsah
Område
Vahemik
Intervalle
Bereich
Αναμενόμενο εύρος α
Intervallo
Diapazons
Intervalas
Område
Zakres
Intervalo
Interval
диапазон
Opseg
Rozsah
Območje
Intervalo
Intervall
Aralığı

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series

Lot: 22509-11

Prod.: Con-Diff Low

ID: 2250911+

Exp.: 30-Jan-2026

Instr.: M-series

Lot: 22509-12

Prod.: Con-Diff Norm

ID: 2250912+

Exp.: 30-Jan-2026

Instr.: M-series

Lot: 22509-13

Prod.: Con-Diff High

ID: 2250913+

Exp.: 3-Feb-2026

1:



BouleCBV2



BouleCBV2



BouleCBV2

2:



2250911+



2250912+



2250913+

3:



26013000001153030000



26013000002153030000



26020300003153030000

4:



10004030800000000000



10004030400000000000



10004029700000000000

5:



22512678500560315325



41518810501180433630



50922929501620547335

6:



08015116152491536630



23630119152841535130



50260121153181534230

7:



03503015040020301808



08606040060060504010



20618125150121006925

8:



15540437700545050970



14140460500655047550



13240610500565033450

9:



11111705288072260928



11112405102043678996



11112941808388910782

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	22509-1K	2026-01-30
Low	22509-11	2026-01-30
Normal	22509-12	2026-01-30
High	22509-13	2026-02-03

SWELAB

Swelab Alfa/Alfa Plus		CONTROL	L
		LOT	22509-11
		Mean ²	± Range ³
RBC	10 ¹² /L	2.26	0.12
MCV	fL	69.2	5.0
RDW	%	14.8	4.0
Hct	%	15.6	2.5
Plt	10 ⁹ /L	80	15
MPV	fL	12.5	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	25.2	1.5
MCHC	g/dL	36.5	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	47.5	7.0
Mid	%	3.6	5.0
Gran	%	48.9	7.0

CONTROL		N
LOT		22509-12
Mean ²	± Range ³	
4.17	0.18	
82.3	5.0	
12.8	4.0	
34.3	3.0	
241	30	
12.7	1.5	
8.6	0.6	
11.9	0.4	
28.5	1.5	
34.7	3.0	
4.2	0.6	
0.3	0.5	
4.1	1.0	
49.4	5.0	
3.5	5.0	
47.1	5.0	

CONTROL		H
LOT		22509-13
Mean ²		± Range ³
5.07		0.22
94.4		5.0
11.8		4.0
47.9		3.5
527		60
13.0		1.5
20.3		1.8
16.2		0.5
32.0		1.5
33.8		3.0
13.0		1.5
0.6		1.0
6.7		2.5
63.7		5.0
3.1		5.0
33.2		5.0

1
bg Трайност на отворен флакон 14 дни (bg)
hr Stabilitet otvorene bočice 14 dni (hr)
cs Stabilita otevřených ampulky 14 dní (cs)
da Åben rør stabilitet 14 dage (da)
et Stabilitus avatud viaalis 14 päevade (et)
fr Stabilité en flacon ouvert 14 jour (fr)
de Stabilität geöffneten Flaschen 14 tage (de)
el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it Stabilità della fiala aperta 14 giorni (it)
lv Atvērt flakonu stabilitāti 14 dienas (lv)
lt Stabilitumas atidarius buteliuką 14 dienas (lt)
no Åpen rør/glass stabilitet 14 dager (no)
pl Trwałość po otwarciu fiolek 14 dni (pl)
pt Estabilidade após abertura do frasco 14 dias (pt)
ro Termenul de valabilitate al fiolei desfăcută 14 zile (ro)
ru Стабильность вскрытого флакона 14 дней (ru)
sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl Stabilita otvoreneje injekčneje liekovky 14 dni (sl)
es Estabilidad de la cápsula abierta 14 días (es)
sv Hållbarhet för öppnad flaska 14 dagar (sv)
tr Açık şişe stabilitesi 14 gün (tr)

2
Sredno
Sredstvo
Průměr
Middelverdi
Keskmine
Moyenne
Mittelwert
Μέση
Media
Nozīmē
Vidurkis
Middelverdi
Średnia
Média
Medie
Среднее
Srednja vrednost
Priemer
Povprečno
Media
Medelvärde
Ortalama

3
Обхват
Raspon
Rozsah
Område
Vahemik
Intervalle
Bereich
Αναμενόμενο εύρος α
Intervallo
Diapazons
Intervalas
Område
Zakres
Intervalo
Interval
диапазон
Opseg
Rozsah
Območje
Intervalo
Intervall
Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa

Lot: 22509-11

Prod.: Con-Diff Low

ID: 2250911+

Exp.: 30-Jan-2026

Instr.: Swelab Alfa

Lot: 22509-12

Prod.: Con-Diff Norm

ID: 2250912+

Exp.: 30-Jan-2026

Instr.: Swelab Alfa

Lot: 22509-13

Prod.: Con-Diff High

ID: 2250913+

Exp.: 3-Feb-2026

