

# Certificate of Analysis

**Product** Boule Micropipette Plastic EDTA 10x100 pcs  
Boule Micropipette Plastic EDTA 1x100 pcs

**Product No** 1070030  
1070039

**Lot No** 251145Q8


**Manufacturing Date** 2025-07-16

**Expiration Date** 2028-06-28

| Parameter                 | Specification                |
|---------------------------|------------------------------|
| Volume 20 µL              | CV ≤ 1,0%                    |
| WBC background*           | ≤ 0,1 x 10 <sup>9</sup> /L   |
| RBC background*           | ≤ 0,01 x 10 <sup>12</sup> /L |
| HGB background            | ≤ 0,2 g/dL                   |
| PLT background            | ≤ 10 x 10 <sup>9</sup> /L    |
| *WBC background for Exigo | ≤ 0,2 x 10 <sup>9</sup> /L   |
| *RBC background for Exigo | ≤ 0,03 x 10 <sup>12</sup> /L |

Actual value meets specification.

Date: 2025-10-02



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

