

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32512-5K	2026-05-05
Low	32512-51	2026-05-05
Normal	32512-52	2026-05-05
High	32512-53	2026-05-07

MEDONIC

CONTROL	L
LOT	32512-51

Medonic M-series/M32		Mean ²	± Range ³
RBC	10 ¹² /L	2.22	0.12
MCV	fL	66.2	7.0
RDW	%	15.9	4.5
Hct	%	14.7	3.0
Plt	10 ⁹ /L	82	15
MPV	fL	11.6	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.8	2.0
MCHC	g/dL	37.4	4.0
Lymph#	10 ⁹ /L	1.4	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	43.0	7.0
Mid	%	6.0	5.0
Gran	%	51.0	7.0

CONTROL	N
LOT	32512-52

	Mean ²	± Range ³
	4.17	0.18
	82.3	7.0
	13.4	4.5
	34.3	3.5
	237	30
	12.3	1.5
	8.3	0.6
	12.1	0.4
	29.0	2.0
	35.3	4.0
	3.9	0.6
	0.5	0.5
	3.9	1.0
	46.5	5.0
	6.4	5.0
	47.1	5.0

CONTROL	H
LOT	32512-53

	Mean ²	± Range ³
	5.10	0.22
	92.4	7.0
	13.6	4.5
	47.1	4.0
	513	60
	12.3	1.5
	20.0	1.8
	16.3	0.5
	32.0	2.0
	34.6	4.0
	12.0	1.5
	1.2	1.0
	6.8	2.5
	60.4	5.0
	5.8	5.0
	33.8	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevrené ampulky 14 dny (cs)
 - da Åben r r stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 p evade (et)
 - fr Stabilit  en flacon ouvert 14 jour (fr)
 - de Stabilit t ge ffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημε ρα (el)
 - it Stabilit  della fiala aperta 14 giorni (it)
 - lv Atv rt flakonu stabilit ti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuk  14 dienas (lt)
 - no  pen r r/glass stabilitet 14 dager (no)
 - pl Trwa o c po otwarciu fiolek 14 Liczba dni (pl)
 - pt Estabilidade ap s abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desf cute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injek nej liekovky 14 dn  (sk)
 - sl Stabilita otvorenej injek nej liekovky 14 dni (sl)
 - es Estabilidad de la c psula abierta 14 dias (es)
 - sv H llbarhet f r  ppnad flaska 14 dagar (sv)
 - tr A ık ŐiŐe stabilitesi 14 g n (tr)

- 2
- Sredno
 - Sredstvo
 - Pr m r
 - Middelv rdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - M Őh
 - Media
 - Noz m 
 - Vidurkis
 - Middelverdi
 - Őrednia
 - M dia
 - Medie
 - Srednee
 - Srednja vrednost
 - Priemer
 - Povpre no
 - Media
 - Medelv rde
 - Ortalama

- 3
- Ob vat
 - Raspon
 - Rozsah
 - Omr de
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμεν μενο  ρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Omr de
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Obmo je
 - Intervalo
 - Intervall
 - Aralıđı

25122-4



[http://www.medonic.se/
support/download](http://www.medonic.se/support/download)

Boule
Boule Medical AB
Domnarvsgatan 4
SE-163 53 Sp nga,
Sweden

For Medonic M-series / M32 with barcode reader





















M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.:	M-series	Instr.:	M-series	Instr.:	M-series
Lot:	32512-51	Lot:	32512-52	Lot:	32512-53
Prod.:	Con-Diff Low	Prod.:	Con-Diff Norm	Prod.:	Con-Diff High
ID:	3251251+	ID:	3251252+	ID:	3251253+
Exp.:	5-May-2026	Exp.:	5-May-2026	Exp.:	7-May-2026

1:	 Boul e CBAV2	 Boul e CBAV2	 Boul e CBAV2
2:	 3251251+	 3251252+	 3251253+
3:	 26050500001153030000	 26050500002153030000	 26050700003153030000
4:	 10004030000000000000	 10004030100000000000	 10004030100000000000
5:	 22212662700550314730	 41718823701210434335	 51022924701630547140
6:	 08215116152482037440	 23730123152902035340	 51360123153202034640
7:	 03303014040020301708	 08306039060050503910	 20018120150121006825
8:	 15945430700605051070	 13445465500645047150	 13645604500585033850
9:	 11113081453535127015	 11110661147769510783	 11112420505338993213

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32512-5K	2026-05-05
Low	32512-51	2026-05-05
Normal	32512-52	2026-05-05
High	32512-53	2026-05-07

SWELAB

		CONTROL	L
		LOT	32512-51
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.24	0.12
MCV	fL	67.6	7.0
RDW	%	15.1	4.5
Hct	%	15.1	3.0
Plt	10 ⁹ /L	80	15
MPV	fL	12.4	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	25.0	2.0
MCHC	g/dL	37.1	4.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	47.4	7.0
Mid	%	3.2	5.0
Gran	%	49.4	7.0

		CONTROL	N
		LOT	32512-52
		Mean ²	± Range ³
		4.17	0.18
		83.5	7.0
		12.4	4.5
		34.8	3.5
		241	30
		12.9	1.5
		8.4	0.6
		12.2	0.4
		29.3	2.0
		35.1	4.0
		4.2	0.6
		0.3	0.5
		3.9	1.0
		50.3	5.0
		3.6	5.0
		46.1	5.0

		CONTROL	H
		LOT	32512-53
		Mean ²	± Range ³
		5.06	0.22
		94.2	7.0
		12.6	4.5
		47.7	4.0
		536	60
		13.2	1.5
		20.1	1.8
		16.3	0.5
		32.2	2.0
		34.2	4.0
		12.8	1.5
		0.6	1.0
		6.7	2.5
		63.6	5.0
		3.2	5.0
		33.2	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dny (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozimė
 - Vidurkis
 - Middelverdi
 - Šrednia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

25122-4



www.swelab.com/
support/download



For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 32512-51
Prod.: Con-Diff Low
ID: 3251251+
Exp.: 5-May-2026

Instr.: Swelab Alfa
Lot: 32512-52
Prod.: Con-Diff Norm
ID: 3251252+
Exp.: 5-May-2026

Instr.: Swelab Alfa
Lot: 32512-53
Prod.: Con-Diff High
ID: 3251253+
Exp.: 7-May-2026

