

# Certificate of Analysis

**Product** SWELAB LUMI-L2 LYSE (500 mL)  
**Product No** 1504516  
**Lot No** 2025101401  
**Manufacturing Date** 2025-10-14  
**Expiration Date** 2027-10-13

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC ( $\times 10^9/L$ )	$\leq 0,2$

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2026-01-28



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

