
Certificate of Analysis

Product SWELAB ALFADILUENT (20 L)
Product No 1504124
Lot No 2601-406
Manufacturing Date 2026-01-19
Expiration Date 2029-01-18

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2026-01-21



Elin Andersson
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

