

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22601-41	2026-06-04
Normal	22601-42	2026-06-04
High	22601-43	2026-06-05

MEDONIC

CONTROL		L	
LOT		22601-41	
		Mean ²	± Range ³
RBC	10 ¹² /L	2.25	0.12
MCV	fL	68.7	5.0
RDW	%	16.2	4.0
Hct	%	15.5	2.5
Plt	10 ⁹ /L	81	15
MPV	fL	11.2	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	25.3	1.5
MCHC	g/dL	36.8	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	43.4	7.0
Mid	%	6.2	5.0
Gran	%	50.4	7.0

CONTROL		N	
LOT		22601-42	
		Mean ²	± Range ³
		4.20	0.18
		79.5	5.0
		14.2	4.0
		33.4	3.0
		234	30
		11.3	1.5
		8.2	0.6
		11.7	0.4
		27.9	1.5
		35.0	3.0
		3.8	0.6
		0.5	0.5
		3.9	1.0
		45.8	5.0
		6.1	5.0
		48.1	5.0

CONTROL		H	
LOT		22601-43	
		Mean ²	± Range ³
		5.14	0.22
		91.2	5.0
		13.3	4.0
		46.9	3.5
		532	60
		11.5	1.5
		20.3	1.8
		16.1	0.5
		31.3	1.5
		34.3	3.0
		12.4	1.5
		1.2	1.0
		6.7	2.5
		61.1	5.0
		5.8	5.0
		33.1	5.0

- ¹
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dana (hr)
 - cs Stabilita otevrené ampulky 14 dní (cs)
 - da Åben ror stabilitet 14 dage (da)
 - et Stabiilsus avatud viaalis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonū stabilitāti 14 dienas (lv)
 - lt Stabilitumas atidarius buteliuka 14 dienas (lt)
 - no Åpen ror/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 L.iczbadni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fioleidestăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- ²
- Sredno
 - Sredstvo
 - Prümër
 - Middelvardi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Mέση
 - Media
 - Nozime
 - Vidurkis
 - Middelverdi
 - Šrednia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Mødelivårde
 - Ortalama

- ³
- Обхват
 - Razpon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - διαпазон
 - Opseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Arahđı

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
 Lot: 22601-41
 Prod.: Con-Diff Low
 ID: 2260141+
 Exp.: 4-Jun-2026

Instr.: M-series
 Lot: 22601-42
 Prod.: Con-Diff Norm
 ID: 2260142+
 Exp.: 4-Jun-2026

Instr.: M-series
 Lot: 22601-43
 Prod.: Con-Diff High
 ID: 2260143+
 Exp.: 5-Jun-2026



Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22601-41	2026-06-04
Normal	22601-42	2026-06-04
High	22601-43	2026-06-05

SWELAB

Swelab Alfa/Alfa Plus	CONTROL LOT	L 22601-41	
		Mean ²	± Range ³
RBC	10 ⁹ /L	2.26	0.12
MCV	fL	70.3	5.0
RDW	%	15.3	4.0
Hct	%	15.9	2.5
Plt	10 ⁹ /L	82	15
MPV	fL	11.7	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	25.2	1.5
MCHC	g/dL	35.8	3.0
Lymph#	10 ⁹ /L	1.7	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	47.9	7.0
Mid	%	3.9	5.0
Gran	%	48.2	7.0

CONTROL LOT	N 22601-42	
	Mean ²	± Range ³
	4.17	0.18
	81.5	5.0
	13.0	4.0
	34.0	3.0
	238	30
	11.9	1.5
	8.2	0.6
	11.7	0.4
	28.1	1.5
	34.4	3.0
	4.1	0.6
	0.3	0.5
	3.8	1.0
	49.7	5.0
	3.9	5.0
	46.4	5.0

CONTROL LOT	H 22601-43	
	Mean ²	± Range ³
	5.11	0.22
	93.2	5.0
	12.3	4.0
	47.6	3.5
	539	60
	12.3	1.5
	20.3	1.8
	16.1	0.5
	31.5	1.5
	33.8	3.0
	13.1	1.5
	0.7	1.0
	6.5	2.5
	64.5	5.0
	3.5	5.0
	32.0	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dana (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 pñevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αλοσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - it Stabilitas atidarius buteliuką 14 dienas (it)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarzeniu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfacute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja rakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Sredno
 - Sredstvo
 - Průměr
 - Middelvaerd
 - Keskmine
 - Moyenne
 - Mittelwert
 - Méση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Srednee
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Ανομενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
 Lot: 22601-41
 Prod.: Con-Diff Low
 ID: 2260141+
 Exp.: 4-Jun-2026

Instr.: Swelab Alfa
 Lot: 22601-42
 Prod.: Con-Diff Norm
 ID: 2260142+
 Exp.: 4-Jun-2026

Instr.: Swelab Alfa
 Lot: 22601-43
 Prod.: Con-Diff High
 ID: 2260143+
 Exp.: 5-Jun-2026

