

# Certificate of Analysis

**Product** MEDONIC M-SERIES LYSE (5 L)  
**Product No** 1504123  
**Lot No** 2601-410  
**Manufacturing Date** 2026-01-26  
**Expiration Date** 2029-01-25

| Test                    | Specification |
|-------------------------|---------------|
| pH                      | 6,80 – 7,20   |
| Osmolality (mOsm/kg)    | 220 – 240     |
| Conductivity (mS/cm)    | 14,00 – 15,00 |
| CBC Performance test*   | Passes test   |
| Particles (WBC channel) | ≤ 0,2         |

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2026-02-02



*Elin Andersson*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

