

# Certificate of Analysis

**Product** BOULE HYPOCHLORITE 2% CLEANER (500 mL)  
**Product No** 1504113  
**Lot No** 2602-412  
**Manufacturing date** 2026-02-02  
**Expiration Date** 2028-02-01


---

<b>Test Hypochlorite</b>	<b>Specification</b>
Hypochlorite concentration	2.0 – 2.4 %
pH	11.5 – 12.1
Blank	PLT < 100 x10 <sup>9</sup> /L RBC < 0.05 x 10 <sup>12</sup> /L

---

Actual values meet specifications.

Date: 2026-02-05



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

