

# Certificate of Analysis

**Product** SWELAB ALFADILUENT, RFID (20 L)  
**Product No** 1504462  
**Lot No** 2602-417  
**Manufacturing Date** 2026-02-09  
**Expiration Date** 2029-02-08

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2026-02-11



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

