

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22602-1K	2026-07-07
Low	22602-11	2026-07-07
Normal	22602-12	2026-07-07
High	22602-13	2026-07-07

Open vial stability 14 days ¹

MEDONIC		CONTROL	L
Medonic M-series/M32		LOT	22602-11
		Mean ²	± Range ³
RBC	10 ⁹ /L	2.22	0.12
MCV	fL	68.4	5.0
RDW	%	15.0	4.0
Hct	%	15.2	2.5
Plt	10 ⁹ /L	82	15
MPV	fL	11.2	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	25.2	1.5
MCHC	g/dL	36.8	3.0
Lymph#	10 ⁹ /L	1.4	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	43.0	7.0
Mid	%	6.0	5.0
Gran	%	51.0	7.0

CONTROL	N
LOT	22602-12
Mean ²	± Range ³
4.19	0.18
84.0	5.0
13.3	4.0
35.2	3.0
236	30
11.5	1.5
8.4	0.6
12.3	0.4
29.4	2.5
34.9	3.0
3.8	0.6
0.5	0.5
4.1	1.0
45.5	5.0
6.5	5.0
48.0	5.0

CONTROL	H
LOT	22602-13
Mean ²	± Range ³
5.11	0.22
92.3	5.0
13.0	4.0
47.2	3.5
526	60
11.7	1.5
20.4	1.8
16.2	0.5
31.7	1.5
34.3	3.0
12.5	1.5
1.2	1.0
6.7	2.5
61.3	5.0
5.7	5.0
33.0	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Intervallo
 - Intervall
 - Aralığı

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
 Lot: 22602-11
 Prod.: Con-Diff Low
 ID: 2260211+
 Exp.: 7-Jul-2026

Instr.: M-series
 Lot: 22602-12
 Prod.: Con-Diff Norm
 ID: 2260212+
 Exp.: 7-Jul-2026

Instr.: M-series
 Lot: 22602-13
 Prod.: Con-Diff High
 ID: 2260213+
 Exp.: 7-Jul-2026



3.

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22602-1K	2026-07-07
Low	22602-11	2026-07-07
Normal	22602-12	2026-07-07
High	22602-13	2026-07-07

Open vial stability 14 days ¹

SWELAB

CONTROL		L	
LOT		22602-11	
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.25	0.12
MCV	fL	69.9	5.0
RDW	%	14.1	4.0
Hct	%	15.7	2.5
Plt	10 ⁹ /L	79	15
MPV	fL	11.8	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	25.3	1.5
MCHC	g/dL	36.3	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	47.7	7.0
Mid	%	3.9	5.0
Gran	%	48.4	7.0

CONTROL		N	
LOT		22602-12	
		Mean ²	± Range ³
		4.20	0.18
		85.6	5.0
		12.1	4.0
		36.0	3.0
		241	30
		12.1	1.5
		8.5	0.6
		12.4	0.4
		29.5	1.5
		34.4	3.0
		4.2	0.6
		0.3	0.5
		4.0	1.0
		49.5	5.0
		3.9	5.0
		46.6	5.0

CONTROL		H	
LOT		22602-13	
		Mean ²	± Range ³
		5.09	0.22
		93.8	5.0
		12.1	4.0
		47.7	3.5
		532	60
		12.3	1.5
		20.3	1.8
		16.1	0.5
		31.6	1.5
		33.8	3.0
		13.0	1.5
		0.7	1.0
		6.6	2.5
		64.2	5.0
		3.5	5.0
		32.3	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud viaalis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčneje liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelværdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Orseg
	Rozsah
	Območje
	Intervalo
	Intervall
	Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
 Lot: 22602-11
 Prod.: Con-Diff Low
 ID: 2260211+
 Exp.: 7-Jul-2026

Instr.: Swelab Alfa
 Lot: 22602-12
 Prod.: Con-Diff Norm
 ID: 2260212+
 Exp.: 7-Jul-2026

Instr.: Swelab Alfa
 Lot: 22602-13
 Prod.: Con-Diff High
 ID: 2260213+
 Exp.: 7-Jul-2026

