

Certificate of Analysis

Product Quintus 5-part Lyse 5L
Product No 1504292
Lot No 2601-405
Manufacturing Date 2026-01-15
Expiration Date 2027-07-14

Test	Specification
pH	2,7 – 3,0
Osmolality (mOsm/kg)	35 – 44
Conductivity ($\mu\text{S}/\text{cm}$)	1117 – 1541
Background WBC ($\times 10^{12}/\text{L}$)	≤ 0.1

Actual values meet specification.

Date: 2026-03-05



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

