

# Certificate of Analysis

**Product** Quintus 5-part Diluent 20L  
**Product No** 1504291  
**Lot No** 2602-416  
**Manufacturing Date** 2026-02-05  
**Expiration Date** 2028-02-04

Test	Specification
pH	6,8 – 7,2
Osmolality (mOsm/kg)	315 – 350
Conductivity (mS/cm)	18,8 – 19,2
Background PLT ( x 10 <sup>9</sup> /L)	≤ 5

Actual values meet specification.

Date: 2026-03-05



*Ulrika Miefalk*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

