

Certificate of Analysis

Product EXIGO LYSE, RFID (1.9 L)
Product No 1504500
Lot No 2603-430
Manufacturing Date 2026-03-11
Expiration Date 2029-03-10

| Test | Specification |
|-----------------------------|---------------|
| pH | 6.95 – 7.25 |
| Osmolality (mOsm/kg) | 195 - 225 |
| Conductivity (mS/cm) | 12.10 – 12.80 |
| CBC Performance * | Passes test |
| Particles (PLT/RBC channel) | ≤ 5 |

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date:2026-03-24



Elin Andersson
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

