


Boule Cal

CAL**IVD**

Calibrator

Lot	
22603-34	2026-06-14

Open vial stability 5 days ¹

Medonic

Parameter ²	RBC 10 ¹² /L	MCV fL	Plt 10 ⁹ /L	MPV fL	WBC 10 ⁹ /L	Hgb g/dL
Medonic M-series / M32	4.17	80.5	234	11.6	8.5	12.0
Range ± ³	0.10	2.0	15	1.5	0.3	0.2

Swelab

Parameter ²	RBC 10 ¹² /L	MCV fL	Plt 10 ⁹ /L	MPV fL	WBC 10 ⁹ /L	Hgb g/dL
Swelab Alfa / Alfa Plus	4.17	81.4	238	12.2	8.5	12.0
Range ± ³	0.10	2.0	15	1.5	0.3	0.2

	1	2	3
bg	Трайност на отворен флакон	5 дни	Обхват
hr	Stabilnost otvorene bočice	5 dani	Raspon
cs	Stabilita otevřené ampulky	5 dní	Rozsah
da	Åben r�r stabilitet	5 dage	Omr�de
et	Stabiilsus avatud viaalis	5 p�evade arv	Vahemik
fr	Stabilit� en flacon ouvert	5 jour	Intervalle
de	Stabilit�t ge�ffneter Flaschen	5 tage	Bereich
el	Διατ�ρηση δειγματος μετ� την αποσφ�ραγιση	5 ημ�ρα	Αναμενόμενο ε�ρος α
it	Stabilit� della fiala aperta	5 giorni	Intervallo
lv	Atv�rt flakonu stabilit�ti	5 dienas	Diapazons
lt	Stabilumas atidarius buteliuka	5 d.	Intervalas
no	�pen r�r/glass stabilitet	5 dager	Omr�de
pl	Trwa�o�c po otworzeniu fiolki	5 Liczba dni	Zakres
pt	Estabilidade ap�s abertura do frasco	5 dias	Intervalo
ro	Termenul de valabilitate al fiolei desfacute	5 zile	Interval
ru	Стабильность вскрытого флакона	5 дней	диапазон
sr	Proizvod stabilan nakon otvaranja pakovanja	5 dana	Opseg
sk	Stabilita otvorenej injek�nej liekovky	5 dn�	Rozsah
sl	Stabilnost odprte stekleni�ke	5 dni	Obmo�je
es	Estabilidad de la c�psula abierta	5 dias	Intervalo
sv	H�llbarhet f�r �ppnad flaska	5 dagar	Intervall
tr	A�ık ŐiŐe stabilitesi	5 g�n	Aralıđı

www.boule.com

 **Boule**
Boule Medical AB
Fagerstagatan 7
SE-163 53 Sp nga,
Sweden

25123-7

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series

Lot: 22603-34

Prod.: Cal

ID: 2260334+

Exp.: 14-Jun-2026



For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa

Lot: 22603-34

Prod.: Cal

ID: 2260334+

Exp.: 14-Jun-2026

