

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32603-5K	2026-08-01
Low	32603-51	2026-08-01
Normal	32603-52	2026-08-05
High	32603-53	2026-08-07

MEDONIC

		CONTROL	L
		LOT	32603-51
Medonic M-series/M32		Mean ²	± Range ³
RBC	10 ¹² /L	2.24	0.12
MCV	fL	67.4	7.0
RDW	%	15.5	4.5
Hct	%	15.1	3.0
Plt	10 ⁹ /L	79	15
MPV	fL	11.4	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	25.0	2.0
MCHC	g/dL	37.1	4.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	43.9	7.0
Mid	%	5.8	5.0
Gran	%	50.3	7.0

	CONTROL	N
	LOT	32603-52
	Mean ²	± Range ³
	4.17	0.18
	82.8	7.0
	13.5	4.5
	34.5	3.5
	235	30
	11.3	1.5
	8.7	0.6
	12.2	0.4
	29.3	2.0
	35.4	4.0
	3.9	0.6
	0.6	0.5
	4.2	1.0
	45.1	5.0
	6.4	5.0
	48.5	5.0

	CONTROL	H
	LOT	32603-53
	Mean ²	± Range ³
	5.08	0.22
	92.2	7.0
	13.9	4.5
	46.8	4.0
	524	60
	11.8	1.5
	20.4	1.8
	16.2	0.5
	31.9	2.0
	34.6	4.0
	12.5	1.5
	1.2	1.0
	6.7	2.5
	61.4	5.0
	5.8	5.0
	32.8	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dni (hr)
 - cs Stabilita otevrené ampulky 14 dny (cs)
 - da Åben r r stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 p evade (et)
 - fr Stabilit  en flacon ouvert 14 jour (fr)
 - de Stabilit t ge ffneter Flaschen 14 tage (de)
 - el Διατ ρηση δειγματος μετ  την αποσφ ραγιση 14 ημερα (el)
 - it Stabilit  della fiala aperta 14 giorni (it)
 - lv Atv rt flakonu stabilit ti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuk  14 dienas (lt)
 - no  pen r r/glass stabilitet 14 dager (no)
 - pl Trwa o c po otwarciu fiolek 14 Liczba dni (pl)
 - pt Estabilidade ap s abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desf cute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekcnej liekovky 14 dni (sk)
 - sl Stabilita otvorenej injekcnej liekovky 14 dni (sl)
 - es Estabilidad de la c psula abierta 14 dias (es)
 - sv H llbarhet f r  ppnad flaska 14 dagar (sv)
 - tr A ık ŐiŐe stabilitesi 14 g n (tr)

- 2
- Sredno
 - Sredstvo
 - Pr m r
 - Middelv rdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - M ση
 - Media
 - Noz m 
 - Vidurkis
 - Middelverdi
 - Őrednia
 - M dia
 - Medie
 - Srednee
 - Srednja vrednost
 - Priemer
 - Povpre no
 - Media
 - Medelv rde
 - Ortalama

- 3
- Ob vat
 - Raspon
 - Rozsah
 - Omr de
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμεν μενο ε ρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Omr de
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Obmo je
 - Intervalo
 - Intervall
 - Aral ı

25122-4



[http://www.medonic.se/
support/download](http://www.medonic.se/support/download)



For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
 Lot: 32603-51
 Prod.: Con-Diff Low
 ID: 3260351+
 Exp.: 1-Aug-2026

Instr.: M-series
 Lot: 32603-52
 Prod.: Con-Diff Norm
 ID: 3260352+
 Exp.: 5-Aug-2026

Instr.: M-series
 Lot: 32603-53
 Prod.: Con-Diff High
 ID: 3260353+
 Exp.: 7-Aug-2026



3.

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32603-5K	2026-08-01
Low	32603-51	2026-08-01
Normal	32603-52	2026-08-05
High	32603-53	2026-08-07

SWELAB

Swelab Alfa/Alfa Plus	UNIT	CONTROL	L
		LOT	32603-51
		Mean ²	± Range ³
RBC	10 ¹² /L	2.27	0.12
MCV	fL	69.1	7.0
RDW	%	14.6	4.5
Hct	%	15.7	3.0
Plt	10 ⁹ /L	77	15
MPV	fL	12.0	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.8	0.3
MCH	pg	25.6	2.0
MCHC	g/dL	36.9	4.0
Lymph#	10 ⁹ /L	1.7	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	48.3	7.0
Mid	%	3.6	5.0
Gran	%	48.1	7.0

CONTROL	N
LOT	32603-52
Mean ²	± Range ³
4.17	0.18
83.8	7.0
12.4	4.5
34.9	3.5
236	30
11.9	1.5
8.6	0.6
12.2	0.4
29.3	2.0
35.0	4.0
4.3	0.6
0.3	0.5
4.0	1.0
49.6	5.0
4.0	5.0
46.4	5.0

CONTROL	H
LOT	32603-53
Mean ²	± Range ³
5.07	0.22
93.3	7.0
13.0	4.5
47.3	4.0
526	60
12.5	1.5
20.3	1.8
16.2	0.5
32.0	2.0
34.2	4.0
13.1	1.5
0.7	1.0
6.5	2.5
64.4	5.0
3.4	5.0
32.2	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevrené ampulky 14 dny (cs)
 - da Åben r r stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 p evade (et)
 - fr Stabilit  en flacon ouvert 14 jour (fr)
 - de Stabilit t ge ffneter Flaschen 14 tage (de)
 - el Διατ ρηση δειγματος μετ  την αποσφραγιστ  14 ημερ  (el)
 - it Stabilit  della fiala aperta 14 giorni (it)
 - lv Atv rt flakonu stabilit ti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuk  14 dienas (lt)
 - no  pen r r/glass stabilitet 14 dager (no)
 - pl Trwa o  po otwarciu fiolek 14 dni (pl)
 - pt Estabilidade ap s abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfacute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekcnej liekovky 14 dni (sk)
 - sl Stabilita otvorenej injekcnej liekovky 14 dni (sl)
 - es Estabilidad de la c psula abierta 14 dias (es)
 - sv H llbarhet f r  ppnad flaska 14 dagar (sv)
 - tr A ık ŐiŐe stabilitesi 14 g n (tr)

- 2
- Средно
 - Sredstvo
 - Pr m r
 - Middelv rdi
 - Keskm ne
 - Moyenne
 - Mittelwert
 - M ση
 - Media
 - Noz m 
 - Vidurkis
 - Middelverdi
 - Őrednia
 - M dia
 - M die
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povpre no
 - Media
 - Medelv rde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Omr de
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εϋρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Omr de
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Obmo je
 - Intervalo
 - Intervall
 - Aral đı

25122-4



www.swelab.com/
support/download



For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 32603-51
Prod.: Con-Diff Low
ID: 3260351+
Exp.: 1-Aug-2026

Instr.: Swelab Alfa
Lot: 32603-52
Prod.: Con-Diff Norm
ID: 3260352+
Exp.: 5-Aug-2026

Instr.: Swelab Alfa
Lot: 32603-53
Prod.: Con-Diff High
ID: 3260353+
Exp.: 7-Aug-2026

