

LOT: 12604-15
EXP: 2026-06-24
Open Vial Stability 5 days

EXIGO

	WBC 10 ⁹ /L	RBC 10 ¹² /L	Hgb g/dL	MCV fL	PLT 10 ⁹ /L
System specific values	8.5	4.18	10.6	76.1	223
Tolerance limits ±	0.3	0.10	0.2	2.0	15



For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 12604-15
Prod.: Vet-Cal
ID: 1260415+
Exp.: 24-Jun-2026

