

# Certificate of Analysis

**Product** BOULE EASYCLEANER (50 mL)  
**Product No** 1504513  
**Lot No.** 2026032001  
**Manufacturing date** 2026-03-20  
**Expiration Date** 2028-03-19

---

Test Hypochlorite	Specification
Hypochlorite concentration	4.50 – <5 %
pH	12.60 – 13.20
Blank	PLT ≤ 10 x10 <sup>9</sup> /L RBC ≤ 0.02 x 10 <sup>12</sup> /L

---

Actual values meet specifications.

Date: 2026-04-15



*Elin Andersson*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

