

# Certificate of Analysis

**Product** Boule Micropipette Plastic EDTA 10x100 pcs  
Boule Micropipette Plastic EDTA 1x100 pcs

**Product No** 1070030  
1070039

**Lot No** 253645D4

**Manufacturing Date** 2026-04-02

**Expiration Date** 2029-02-28

Parameter	Specification
Volume 20 $\mu$ L	CV $\leq$ 1,0%
WBC background*	$\leq$ 0,1 x 10 <sup>9</sup> /L
RBC background*	$\leq$ 0,01 x 10 <sup>12</sup> /L
HGB background	$\leq$ 0,2 g/dL
PLT background	$\leq$ 10 x 10 <sup>9</sup> /L
*WBC background for Exigo	$\leq$ 0,2 x 10 <sup>9</sup> /L
*RBC background for Exigo	$\leq$ 0,03 x 10 <sup>12</sup> /L

Actual value meets specification.

Date: 2026-04-24



*Ulrika Miefalk*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

