

# Certificate of Analysis

<b>Product</b>	SWELAB ALFALYSE, RFID (5 L)
<b>Product No</b>	1504463
<b>Lot No</b>	2604-440
<b>Manufacturing Date</b>	2026-04-15
<b>Expiration Date</b>	2029-04-14

<b>Test</b>	<b>Specification</b>
pH	7,80 – 8,30
Osmolality (mOsm/kg)	295 – 330
Conductivity (mS/cm)	17,75 – 18,25
CBC Performance *	Passes test
Particles (WBC channel)	≤ 0,2

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2026-04-23



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

