

---

# Certificate of Analysis

**Product** SWELAB LUMI-L2 LYSE (500 mL), 1504516  
**Lot No** 2026022801  
**Manufacturing Date** 2026-02-28  
**Expiration Date** 2028-02-27


---

<b>Test</b>	<b>Specification</b>
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2026-05-20



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

