

Certificate of Analysis

Product SWELAB LUMI-D DILUENT (20L), 1504514
Lot No 2026030901
Manufacturing Date 2026-03-09
Expiration Date 2028-03-08

| Test | Specification |
|---|---------------|
| pH | 7,1 – 7,5 |
| Osmolality (mOsm/kg) | 290 – 310 |
| Conductivity (mS/cm) | 17,3 – 18,3 |
| Particles PLT/RBC channel ($\times 10^9/L$) | ≤ 5 |

Actual values meet specification.

Date: 2026-05-18



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

