


Certificate of Analysis

Product	Quintus 5-part Diluent 20L
Product No	1504291
Lot No	2605-446
Manufacturing Date	2026-05-06
Expiration Date	2028-05-05

Test	Specification
pH	6,8 – 7,2
Osmolality (mOsm/kg)	315 – 350
Conductivity (mS/cm)	18,8 – 19,2
Background PLT (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2026-05-08



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

