

# Certificate of Analysis


**Product** MEDONIC M51-L1 LYSE (200 mL), 1504511  
**Lot No** 2026031801  
**Manufacturing Date** 2026-03-18  
**Expiration Date** 2028-03-17

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2026-06-15



*Ulrika Miefalk*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

