

Certificate of Analysis

Product EXIGO EOS REAGENT (1.9 L)
Product No 1504278
Lot No 2606-466
Manufacturing Date 2026-06-23
Expiration Date 2028-06-22

| Test | Specification |
|----------------------|---------------|
| pH | 8.70 – 9.30 |
| Osmolality (mOsm/kg) | 125 - 145 |
| Conductivity (mS/cm) | 6.80 – 7.60 |
| EOS Performance * | Passes test |
| Blank EOS | ≤ 0.1 |

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory.

Actual values meet specifications.

Date: 2026-07-01



Elin Andersson
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

