

Certificate of Analysis


Product SWELAB LUMI-L1 LYSE (200 mL), 1504515
Lot No 2026022702
Manufacturing Date 2026-02-27
Expiration Date 2028-02-26

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2

*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2026-08-28



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

