


Certificate of Analysis

Product MEDONIC M51-D DILUENT (20L), 1504510
Lot No 2026051902
Manufacturing Date 2026-05-19
Expiration Date 2028-05-18

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel ($\times 10^9/L$)	≤ 5

Actual values meet specification.

Date: 2026-08-28



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

