


# Certificate of Analysis

**Product** MEDONIC M51-D DILUENT (20L), 1504510  
**Lot No** 2026052201  
**Manufacturing Date** 2026-05-22  
**Expiration Date** 2028-05-21

| Test                                             | Specification |
|--------------------------------------------------|---------------|
| pH                                               | 7,1 – 7,5     |
| Osmolality (mOsm/kg)                             | 290 – 310     |
| Conductivity (mS/cm)                             | 17,3 – 18,3   |
| Particles PLT/RBC channel (x 10 <sup>9</sup> /L) | ≤ 5           |

Actual values meet specification.

Date: 2026-08-28



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.