

Boule Con-5Diff A1

CONTROL

IVD

Tri-Level	72607-1K	2026/11/05
Low	72607-11	2026/11/05
Normal	72607-12	2026/11/05
High	72607-13	2026/11/05

Open vial stability: 14 days ¹

Medonic M51 / Swelab Lumi

Parameter ²	Unit ³	Low		Normal		High	
		Mean ⁴	± Range ⁵	Mean ⁴	± Range ⁵	Mean ⁴	± Range ⁵
WBC	10 ⁹ /L	3.52	± 0.50	8.39	± 1.00	19.31	± 2.50
Neu%	%	49.9	16.0	55.8	8.0	64.5	8.0
Lym%	%	40.8	21.0	32.1	8.0	21.8	12.0
Mon%	%	5.4	5.0	5.9	5.0	5.9	6.0
Eos%	%	4.1	5.0	6.4	6.0	7.9	8.0
Bas%	%	2.4	10.0	1.9	10.0	1.3	10.0
Neu#	10 ⁹ /L	1.77	0.50	4.70	0.70	12.56	2.00
Lym#	10 ⁹ /L	1.45	0.70	2.70	0.80	4.24	2.00
Mon#	10 ⁹ /L	0.19	0.30	0.50	0.50	1.14	1.10
Eos#	10 ⁹ /L	0.15	0.20	0.54	0.60	1.55	1.40
Bas#	10 ⁹ /L	0.09	0.40	0.16	0.80	0.26	1.80
RBC	10 ¹² /L	2.40	0.24	4.55	0.30	5.51	0.50
HGB	g/L	63	6	135	6	177	8
HCT	%	19.2	2.0	41.3	3.0	52.9	4.0
MCV	fL	80.0	5.0	90.8	5.0	96.0	6.0
MCH	pg	26.3	2.5	29.7	2.5	32.1	2.5
MCHC	g/L	328	30	327	30	335	30
RDW-CV	%	16.7	3.0	14.8	3.0	14.5	3.0
PLT	10 ⁹ /L	60	20	256	40	470	60
MPV	fL	8.0	3.0	7.5	3.0	7.5	3.0

No mean values and ranges are provided for research parameters-RDW-SD, PDW, PCT, P-LCR, and P-LCC. User may establish these based on their usage.



1	2	3	4	5
bg Трайност на отворен флакон: 14 дни	Параметър	Единица	Средно	Обхват
cs Stabilita otevřené ampulky: 14 dní	Parametr	Jednotka	Průměr	Rozsah
da Åben rør stabilitet: 14 dage	Parameter	Enhed	Middelværdi	Område
de Stabilität geöffnete Flaschen: 14 Tage	Parameter	Einheit	Mittelwert	Bereich
el Διατήρηση δείγματος μετά την αποσφράγιση: 14 ημέρα	Παράμετρος	Μονάδα	Μέση	Αναμενόμενο εύρος α
es Estabilidad de la cápsula abierta: 14 días	Parámetro	Unidad	Media	Intervalo
et Stabiilsus avatud vialis: 14 päevade arv	Parameeter	Ühik	Keskmine	Vahemik
fr Stabilité en flacon ouvert: 14 jour	Paramètre	Unité	Moyenne	Intervalle
hr Stabilnost otvorene bočice: 14 dani	Parametar	Jedinica	Sredstvo	Raspon
it Stabilità della fiala aperta: 14 giorni	Parametro	Unità	Media	Intervallo
lt Stabilumas atidarius buteliuką: 14 dienų	Parametras	Vienetas	Vidurkis	Intervalas
lv Atvērt flakonu stabilitāti: 14 dienas	Parametrs	Vienība	Nozīmē	Diapazons
no Åpen rør/glass stabilitet: 14 dager	Parameter	Enhhet	Middelverdi	Område
pl Trwałość po otwarciu fioleki: 14 Liczba dni	Kryterium	Jednostka	Średnia	Zakres
pt Estabilidade após abertura do frasco: 14 dias	Parâmetro	Unidade	Média	Intervalo
ro Termenul de valabilitate al fiolei desfăcută: 14 zile	Parametru	Unitate	Medie	Interval
ru Стабильность вскрытого флакона: 14 дней	Параметр	Единица измерения	Среднее	диапазон
sk Stabilita otvorenej injekčnej liekovky: 14 dní	Parameter	Jednotka	Priemer	Rozsah
sl Stabilnost odprte stekleničke: 14 dni	Parameter	Enota	Povprečno	Območje
sr Proizvod stabilan nakon otvaranja pakovanja: 14 dana	Parametar	Jedinica	Srednja vrednost	Opseg
sv Hållbarhet för öppnad flaska: 14 dagar	Parameter	Enhet	Medelvärde	Intervall
tr Açık şişe stabilitesi gün: 14 gün	Parametre	Birim	Ortalama	Aralığı



www.boule.com

Boule
Boule Medical AB
Fagerstagatan 7
SE-163 53 Spånga,
Sweden

31340-14