

LOT: 12608-14
EXP: 2026-11-03
Open Vial Stability 5 days

EXIGO

	WBC 10 ⁹ /L	RBC 10 ¹² /L	Hgb g/dL	MCV fL	PLT 10 ⁹ /L
System specific values	8.4	4.12	10.7	77.7	226
Tolerance limits $\pm$	0.3	0.10	0.2	2.0	15



For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo

Lot: 12608-14

Prod.: Vet-Cal

ID: 1260814+

Exp.: 3-Nov-2026

1:



2:



3:



4:



5:



6:



7:



8:



9:

