

# Certificate of Analysis

**Product** BOULE EASYCLEANER (50 mL)  
**Product No** 1504513  
**Lot No.** 2026081302  
  
**Manufacturing date** 2026-08-13  
**Expiration Date** 2028-08-12

| Test Hypochlorite          | Specification                                                    |
|----------------------------|------------------------------------------------------------------|
| Hypochlorite concentration | 4.50 – <5 %                                                      |
| pH                         | 12.60 – 13.20                                                    |
| Blank                      | PLT ≤ 10 x10 <sup>9</sup> /L<br>RBC ≤ 0.02 x 10 <sup>12</sup> /L |

Actual values meet specifications.

Date: 2026-09-01



*Ulrika Miefalk*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

