


# Certificate of Analysis

**Product** H50V LYSE, RFID  
**Product No** 1440110 (200 mL)  
**Lot No** 26040585M3  
**Manufacturing Date** 2026-04-15  
**Expiration Date** 2028-04-14

| Test                             | Specification                                                                        |
|----------------------------------|--------------------------------------------------------------------------------------|
| Appearance                       | Clear liquid, no precipitate particles or flocs                                      |
| pH                               | 2.00 – 4.50                                                                          |
| Absorbance $\lambda_{\max}$ (nm) | 530 – 550                                                                            |
| Blank                            | WBC $\leq 0.3 \times 10^9$ /L HGB $\leq 2$ g/L                                       |
| WBC                              | a) Has two population peaks<br>b) Meet the peak form and position mark requirements. |

Actual values meet specifications.

Date: 2026-09-10



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

