


Certificate of Analysis

Product H50V Probe Cleaner
Product No 1440111 (100 mL)
Lot No 26040854M2
Manufacturing Date 2026-04-20
Expiration Date 2027-04-19

Test	Specification
Appearance	Transparent liquid without precipitation, particle or floccules
pH	12,00 - 14,00
Blank	WBC $\leq 0.5 \times 10^9$ /L RBC $\leq 0.1 \times 10^{12}$ /L HGB ≤ 5 g/L PLT $\leq 20 \times 10^9$ /L

Actual values meet specifications.

Date: 2026-09-10



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

