

Certificate of Analysis

Product MEDONIC M-SERIES DILUENT (20 L)
Product No 1504122
Lot No 2608-487
Manufacturing Date 2026-08-31
Expiration Date 2029-08-30

Test	Specification
pH	6,90 – 7,10
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,50 – 18,00
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2026-09-02



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

