

Certificate of Analysis

Product	SWELAB ALFADILUENT (20 L)
Product No	1504124
Lot No	2608-490
Manufacturing Date	2026-09-14
Expiration Date	2029-09-13

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2026-09-18



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

