

Boule Cal

CAL

IVD

Calibrator	Lot	
	22609-34	2026-12-13

Open vial stability 5 days ¹

Medonic

Parameter ²	RBC 10 ¹² /L	MCV fL	Plt 10 ⁹ /L	MPV fL	WBC 10 ⁹ /L	Hgb g/dL
Medonic M-series / M32	4.12	79.2	228	11.5	8.5	11.7
Range ± ³	0.10	2.0	15	1.5	0.3	0.2

Swelab

Parameter ²	RBC 10 ¹² /L	MCV fL	Plt 10 ⁹ /L	MPV fL	WBC 10 ⁹ /L	Hgb g/dL
Swelab Alfa / Alfa Plus	4.12	79.9	233	12.1	8.5	11.7
Range ± ³	0.10	2.0	15	1.5	0.3	0.2

	1
bg	Трайност на отворен флакон
hr	Stabilnost otvorene bočice
cs	Stabilita otevřené ampulky
da	Åben rør stabilitet
et	Stabiilsus avatud viaalis
fr	Stabilité en flacon ouvert
de	Stabilität geöffneter Flaschen
el	Διατήρηση δείγματος μετά την αποσφράγιση
it	Stabilità della fiala aperta
lv	Atvērt flakonu stabilitāti
lt	Stabilumas atidarius buteliuką
no	Åpen rør/glass stabilitet
pl	Trwałość po otworzeniu fiolki
pt	Estabilidade após abertura do frasco
ro	Termenul de valabilitate al fiolei desfăcute
ru	Стабильность вскрытого флакона
sr	Proizvod stabilan nakon otvaranja pakovanja
sk	Stabilita otvorenej injekčnej liekovky
sl	Stabilnost odprte stekleničke
es	Estabilidad de la cápsula abierta
sv	Hållbarhet för öppnad flaska
tr	Açık şişe stabilitesi

	2	3
5 дни	Параметър	Обхват
5 dani	Parametar	Raspon
5 dny	Parametr	Rozsah
5 dage	Parameter	Område
5 päevade arv	Parameeter	Vahemik
5 jour	Paramètre	Intervalle
5 tage	Parameter	Bereich
5 ημέρα	Παράμετρος	Αναμενόμενο εύρος α
5 giorni	Parametro	Intervallo
5 dienas	Parametrs	Diapazons
5 d.	Parametras	Intervalas
5 dager	Parameter	Område
5 Liczba dni	Kryterium	Zakres
5 dias	Parâmetro	Intervalo
5 zile	Parametru	Interval
5 дней	Параметр	диапазон
5 dana	Parametar	Opseg
5 dní	Parameter	Rozsah
5 dni	Parameter	Območje
5 dias	Parâmetro	Intervalo
5 dagar	Parameter	Intervall
5 gün	Parametre	Aralığı



www.boule.com

 **Boule**

Boule Medical AB
Fagerstagatan 7
SE-163 53 Spånga,
Sweden

25123-7

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series

Lot: 22609-34

Prod.: Cal

ID: 2260934+

Exp.: 13-Dec-2026



For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa

Lot: 22609-34

Prod.: Cal

ID: 2260934+

Exp.: 13-Dec-2026

