

Certificate of Analysis

Product MEDONIC M51-L2 LYSE (500 mL), 1504512

Lot No 2026071001

Manufacturing Date 2026-07-10

Expiration Date 2028-07-09

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC ($\times 10^9/L$)	$\leq 0,2$

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2026-10-08



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

