


Certificate of Analysis

Product SWELAB LUMI-D DILUENT (20L), 1504514
Lot No 2026081401
Manufacturing Date 2026-08-14
Expiration Date 2028-08-13

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2026-10-20



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

