

Certificate of Analysis

Product EXIGO DILUENT
Product No 1504328 (10 L)
Lot No 2608-494

Manufacturing Date 2026-10-06

Expiration Date 2029-10-05

Test	Specification
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5

Actual values meet specifications.

Date: 2026-10-12



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

